

TERRA AIR SERVICES

Tunable Diode Laser Summary Report

Customer: Tosco SFAR
Address: 1380 San Pablo Avenue
Rodeo, CA 94572-1299

Monitoring Period

From: 12/01/98 00:00:01 PST
To: 12/31/98 23:59:59 PST



2950 N. Loop W., Ste. 1050
Houston, Texas 77092
Phone (713) 956-2165
Fax (713) 956-2185

North Fenceline

Compound	Average Concentration (ppm)				Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#				
Ammonia	0.000	0.000	0.000	100	0	3	0	744	84.21	15.79	
Hydrogen sulfide	0.000	0.000	0.000	15	0	0.03	0	744	84.22	15.78	

South Fenceline

Compound	Average Concentration (ppm)				Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#				
Ammonia	0.000	0.000	0.000	100	0	3	0	744	82.57	17.43	
Hydrogen sulfide	0.000	0.000	0.000	15	0	0.03	0	744	83.01	16.99	